

# BARRICADE™ COIL SYSTEM

Patient Name: \_\_\_\_\_

Date of Procedure: \_\_\_\_\_

Treating Physician: \_\_\_\_\_

Physician Phone Number: \_\_\_\_\_

Hospital: \_\_\_\_\_

**BLOCKADE MEDICAL**

18 TECHNOLOGY DRIVE #169 | IRVINE CA 92618

p: 949.788.1443 | f: 949.788.1444

**WWW.BLOCKADEMEDICAL.COM**

MKTG-033 Rev. B

# BARRICADE™ COIL SYSTEM

## MRI Compatibility<sup>1</sup>:



The Barricade Coil System implant materials have been determined to be MR Conditional\*. A patient with this device can be scanned safely under the following conditions:

- Static Magnetic field of **3-Tesla**
- Maximum spatial gradient field of **4000 Gauss/cm (40 Tesla/m) or less**
- Maximum MR system reported, whole body averaged specific absorption rate **(SAR) of 2.9 W/kg for 15 minutes of scanning**

Under the scan conditions defined above, the Barricade Coil System is expected to produce a maximum temperature rise of 1.9°C after 15 minutes of continuous scanning.

---

\* Non-Clinical testing has demonstrated that the Barricade Coil System is MR Conditional.

<sup>1</sup> Test data on file at Blockade Medical, LLC